

Certificate of AnalysisProduct : **ALUMINIUM OXIDE NEUTRAL**Batch No. **2137450114**

Description : White powder without any odour.

Grade : **SQ****QUALIGENS****Mfg. Date : January 2014**

Application : For analysis

Expiry Phrase : Use within 5yrs of opening subject to
adhering to the storage conditions

SPECIFICATIONS	UNIT	STANDARDS	RESULTS
Description	----	White powder without any odour.	White powder without any odour.
Identify	----	Passes test	Passes test
Sieve analysis between 80# & 230#	%	90 min	98.1
Percolation time	minutes	5 to 15 minutes.	5 minutes 6 sec.
Adsorptivity	----	Satisfactory	Satisfactory.
pH of 5% w/v aq. Soln.	----	5.5 – 7	6.0

Additional Information :

The product has been tested by the Quality Control Laboratory of Thermo Fisher Scientific and conforms to the above specifications



Geetha Karthikeyan
Q.A. Manager

Any queries please phone : +91- 022 66803004